

UPC MARATHON

A NITREX COMPANY

OXYGEN PROBE CERTIFICATION

SUITABLE FOR ANY MAKE OF OXYGEN PROBE

UPC-Marathon's probe certification program is designed to maximize the life of a probe and ensure the most accurate readings possible. All probes are tested in a certified atmosphere against a probe standard and must meet deviation criteria.

OUR PROGRAM INCLUDES THE FOLLOWING:

- Mechanical Inspection
- Millivolt (sensor output) measurement in a certified atmosphere
- Probe thermocouple verification
- Impedance Testing
- Recovery Testing
- Leak Testing/Reference air flow verification

If a probe should fail testing, UPC-Marathon recommends repair or replacement.



OXYGEN PROBE

- COMPREHENSIVE INSPECTION
- LEAK, RECOVERY & IMPEDANCE TESTS
- EXPERT VERIFICATION



Eco-friendly technology

OUR LARGE CAPACITY TEST FURNACES HELP KEEP TURNAROUND TIMES FAST.


UPC-Marathon offers instrumentation and accessories that can automatically perform some of these tests in-situ.



Reference Air



Protherm 470



Protherm 710

USA

+1 414 462 8200
sales.na@upc-marathon.com

CANADA

+1 514 335 7191
sales.na@upc-marathon.com

CHINA

+86 21 3463 0376
sales@mmichina.cn

FRANCE

+33 3 81 48 37 37
sales.fr@upc-marathon.com

GERMANY

+49 7161 94888 0
sales.de@upc-marathon.com

POLAND

+48 32 296 66 30
sales.pl@upc-marathon.com

UPC MARATHON

A NITREX COMPANY

EVALUATION TECHNIQUES ENSURE THE PERFORMANCE OF YOUR PROBE

MECHANICAL INSPECTION

Each probe is carefully inspected for condition by a skilled UPC technician. Any indications of soot, contamination, fatigue, chemical attack and any other mechanical problem are noted and recommendations are provided based on the "as found" condition of the probe.

MILLIVOLT MEASUREMENT (SENSOR OUTPUT)

Probe millivolt output is measured in a certified atmosphere and compared to a reference standard. Certification contains reference standard reading, allowable deviation, actual reading, and pass/fail result.

PROBE THERMOCOUPLE VERIFICATION

Probes are tested at 1750°F (954°C) against a certified standard.

Certification contains reference standard reading, allowable deviation, actual reading, and pass/fail result.

INDEPENDENCE TEST

Impedance testing is a long-accepted, standard method for the indication of probe condition. This test clearly assesses the health of the probe's electrode contact.

As a probe ages, the millivolt output and recovery time are affected by minute soot particulates in the electrode interface, or by mechanical and contamination issues with the electrode interface. For this test, the millivolt output of the sensor is compared to the reading taken prior to the test.

RECOVERY TEST

With time and usage, it's possible for the zirconia electrolyte of the probe to destabilize and for the electrode contact to deteriorate.

Recovery time of the millivolt reading before and after the impedance test provides an excellent indication of these conditions.

LEAK TESTING / REFERENCE AIR VERIFICATION

During the test, the sensor is placed in the furnace atmosphere and the reference air supply is disconnected. If there's a leakage in the probe electrolyte, the millivolt output would drop beyond the tolerance limit.

OXYGEN PROBE

- COMPREHENSIVE INSPECTION
- LEAK, RECOVERY & IMPEDANCE TESTS
- EXPERT VERIFICATION



Eco-friendly technology



Probe Thermocouple Verification

Main Menu			
Impedance		Burnoff	
Date	5/29/14	Date	5/29/14
Time	12:42	Time	12:42
Impedance(ohm)	3.1K	Temperature	1749
Recovery Time	0:04	Mesured mV	1147
		Recovery Time	0:58

UPC-MARATHON			
PROBE MANUFACTURING AND TEST CERTIFICATE			
Description	20" (500mm) NITROCARB "C" T/C 1/4" CPI		
Serial Number	14031562201	T/C Type	K
Part Number	AA620-K-19	Furnace	2
Test Number	1404021050	Furnace Port	4
Time of Test	10:50	Reference Temp	1750°F
Initials	GSTAFF	Date of Test	04/02/14
	Measurement	Standard	Tolerance
Sensor Temp	1725 °F	1750 °F	± 25 °F
mV Reading before impedance	84.3 mV	80.2 mV	± 5.0 mV
Impedance	2.9 KΩ	± 3 KΩ	PASSED
mV Reading after impedance	84.2 mV	± 3 mV deviation	PASSED
Impedance Recovery Time	1.0 sec	± 2 Sec	PASSED
Ref air shut-off, leak test mV	85.4 mV	± -3 mV deviation	PASSED
RESULTS CERTIFIED BY: _____			

Probe Test Certification

USA

+1 414 462 8200
sales.us@upc-marathon.com

CANADA

+1 514 335 7191
sales.ca@upc-marathon.com

CHINA

+86 21 3463 0376
sales@mmichina.cn

FRANCE

+33 3 81 48 37 37
sales.fr@upc-marathon.com

GERMANY

+49 7161 94888 0
sales.de@upc-marathon.com

POLAND

+48 32 296 66 30
sales.pl@upc-marathon.com